


CALIBRATION POTENTIOMETERS ON THIS SIDE

FIELD NOS.
A=2
B=1
C=3

NOTE:
POSITIVE OUTPUT

ION CHAMBER, 4 FIELD
FOR 14" x 36" CASSETTE, LOAD L.H.
ION CHAMBER THICKNESS: 0.4" (10.2 mm)
PRE-AMP BOARD: 61135G

MFR. - AID INC. (MELROSE PARK, IL.)
MFR. P/N - 1C x 184

X.X ±.030		TOLERANCE UNLESS OTHERWISE SPECIFIED					
X.XX ±.015		 SUMMIT INDUSTRIES, INC. 2901 W. LAWRENCE AVE, CHICAGO, IL TEL 773-509-6273 FAX 773-588-6820					
X.XXX ±.005							
1/X ±1/64							
ANGLES ±1/2°							
TITLE AEC, ION CHAMBER, 36" L.H. (IC x 184)							
NEXT ASSY	MATERIAL						
-	SEE ABOVE						
SCALE	FINISH						
NONE	NONE						
DWG BY	DATE	APPROVED	DATE	SIZE	SHEET	PART NO.	REV
RF	5/28/02	CJS	5/30/02	A	1 OF 1	00892	B

3463	ADDED DIM 1.63, AND NOTE	LG	09/24/03
2614	RELEASED	RF	5/28/02
LTR	REVISION	BY	DATE